

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/25/2021

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:47
INTERNAL STANDARD	VERIFIED	17:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:48
BLANK TEST	0.00	17:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:49
BLANK TEST	0.00	17:50

PO Steve Bunio #289

