

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/28/2021

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:42
INTERNAL STANDARD	VERIFIED	14:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:43
BLANK TEST	0.00	14:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:44
BLANK TEST	0.00	14:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:45
BLANK TEST	0.00	14:46

Handwritten signature and initials

