

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/01/2021

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:19
INTERNAL STANDARD	VERIFIED	14:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:21
BLANK TEST	0.00	14:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:22
BLANK TEST	0.00	14:23



John D Bateholts #714