

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/07/2021

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:11
INTERNAL STANDARD	VERIFIED	15:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:11
BLANK TEST	0.00	15:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:13
BLANK TEST	0.00	15:14

Steven Bunio #983

