

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/14/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:20
INTERNAL STANDARD	VERIFIED	20:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:22
BLANK TEST	0.00	20:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:23
BLANK TEST	0.00	20:24

*PC/MCA*

