

SOLUTION CHANGE

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 12/01/2021

OPERATOR NAME:
 RYAN C SCHROECK
 Permit #: 41316
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:15
INTERNAL STANDARD	VERIFIED	20:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:15
BLANK TEST	0.00	20:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:16
BLANK TEST	0.00	20:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:17
BLANK TEST	0.00	20:18

RC

