

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 12/24/2021

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:07
INTERNAL STANDARD	VERIFIED	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:09
BLANK TEST	0.00	15:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:10
BLANK TEST	0.00	15:11

