

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/15/2021

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:48
INTERNAL STANDARD	VERIFIED	14:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:49
BLANK TEST	0.00	14:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:50
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:51
BLANK TEST	0.00	14:52

