

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 01/31/2021

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:38
INTERNAL STANDARD	VERIFIED	11:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:39
BLANK TEST	0.00	11:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:40
BLANK TEST	0.00	11:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:41
BLANK TEST	0.00	11:42

*Ch. B. Geisler* #248

