

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 02/28/2021

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:52
INTERNAL STANDARD	VERIFIED	08:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:52
BLANK TEST	0.00	08:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:54
BLANK TEST	0.00	08:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:55
BLANK TEST	0.00	08:56

Charles B Geisler

