

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/16/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:07
INTERNAL STANDARD	VERIFIED	19:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:10
BLANK TEST	0.00	19:11

*207/1764*

