

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 09/08/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:24
INTERNAL STANDARD	VERIFIED	23:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:25
BLANK TEST	0.00	23:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:26
BLANK TEST	0.00	23:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:27
BLANK TEST	0.00	23:28

*RS/SW*

