

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/28/2021

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:03
INTERNAL STANDARD	VERIFIED	15:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:06
BLANK TEST	0.00	15:07

