

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 11/01/2021

OPERATOR NAME:  
JOHN D BATEHOLTS  
Permit #: 43007  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:07
INTERNAL STANDARD	VERIFIED	16:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:08
BLANK TEST	0.00	16:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:09
BLANK TEST	0.00	16:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:10
BLANK TEST	0.00	16:11



*R. [Signature]* #714