

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/14/2021

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:25
INTERNAL STANDARD	VERIFIED	19:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:26
BLANK TEST	0.00	19:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:27
BLANK TEST	0.00	19:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29

RS/S

