

SOLUTION CHANGE

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 12/01/2021

OPERATOR NAME:
 RYAN C SCHROECK
 Permit #: 41316
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:23
INTERNAL STANDARD	VERIFIED	21:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:23
BLANK TEST	0.00	21:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:24
BLANK TEST	0.00	21:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:25
BLANK TEST	0.00	21:26

RS

