

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 12/10/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:14
INTERNAL STANDARD	VERIFIED	19:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:17
BLANK TEST	0.00	19:18

*RC/S*

