

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/22/2021

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:56
INTERNAL STANDARD	VERIFIED	13:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:57
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:58
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:59
BLANK TEST	0.00	14:00

JML 2021

