

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/31/2021

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:14
INTERNAL STANDARD	VERIFIED	15:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:14
BLANK TEST	0.00	15:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:16
BLANK TEST	0.00	15:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:17
BLANK TEST	0.00	15:18



10 Steve Bunio 789