

SOLUTION CHANGE

-----  
 WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 02/28/2021

OPERATOR NAME:  
 CHARLES B GEISLER  
 Permit #: 33757  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:23
INTERNAL STANDARD	VERIFIED	11:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:23
BLANK TEST	0.00	11:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:25
BLANK TEST	0.00	11:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:26
BLANK TEST	0.00	11:27

*Charles B Geisler*

