

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/22/2021

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:09
INTERNAL STANDARD	VERIFIED	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:10
BLANK TEST	0.00	09:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:11
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:12
BLANK TEST	0.00	09:13

P.O. *[Signature]* #301

03/22/2021

