

SOLUTION CHANGE




WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 03/31/2021

OPERATOR NAME:  
 CHARLES B GEISLER  
 Permit #: 33757  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:56
INTERNAL STANDARD	VERIFIED	14:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:56
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59



*Handwritten signature: C. B. Geisler #210*