

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/19/2021

OPERATOR NAME:  
JOHN D BATEHOLTS  
Permit #: 43007  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:20
INTERNAL STANDARD	VERIFIED	14:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:21
BLANK TEST	0.00	14:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:23
BLANK TEST	0.00	14:24



*John D Bateholts #714*