

SUPERVISOR TEST

WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 04/25/2021

OPERATOR NAME:  
 CHARLES B GEISLER  
 Permit #: 33757  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:42
BLANK TEST	0.00	13:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:44
BLANK TEST	0.00	13:45

*Charles B Geisler*

