

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/01/2021

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:14
INTERNAL STANDARD	VERIFIED	23:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:15
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:16
BLANK TEST	0.00	23:17

RS/S

