

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/08/2021

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:08
INTERNAL STANDARD	VERIFIED	12:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:09
BLANK TEST	0.00	12:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:10
BLANK TEST	0.00	12:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:11
BLANK TEST	0.00	12:12

*[Handwritten signatures and initials]*

