

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/02/2021

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:01
INTERNAL STANDARD	VERIFIED	14:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:04
BLANK TEST	0.00	14:04

*P.O. [Signature] #301
7/2/2021*

