

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/09/2021

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57



PO M. L. Q. #757