

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/09/2021

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:49
INTERNAL STANDARD	VERIFIED	13:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:49
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:50
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:51
BLANK TEST	0.00	13:52

*gpt. C.B. Geisler #248*

