

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/16/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:17
INTERNAL STANDARD	VERIFIED	21:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:17
BLANK TEST	0.00	21:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:18
BLANK TEST	0.00	21:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:20
BLANK TEST	0.00	21:20

*RS/S*

