

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/16/2021

OPERATOR NAME:  
JOHN D BATEHOLTS  
Permit #: 43007  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:40
INTERNAL STANDARD	VERIFIED	09:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:40
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:42
BLANK TEST	0.00	09:43



*John D. Bateholts* 09/16/21