

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/25/2021

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:46
INTERNAL STANDARD	VERIFIED	16:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:46
BLANK TEST	0.00	16:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:47
BLANK TEST	0.00	16:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:48
BLANK TEST	0.00	16:49

*PO Steve Bunio #7827*

