

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 10/17/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:50
INTERNAL STANDARD	VERIFIED	01:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:52
BLANK TEST	0.00	01:53

*RS 724*

