

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/24/2021

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:03
INTERNAL STANDARD	VERIFIED	09:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:04
BLANK TEST	0.00	09:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:06
BLANK TEST	0.00	09:07

