

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/28/2021

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:28
INTERNAL STANDARD	VERIFIED	15:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:29
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:30
BLANK TEST	0.00	15:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:31
BLANK TEST	0.00	15:32

Handwritten signature SRS

