

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 11/07/2021

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:37
INTERNAL STANDARD	VERIFIED	14:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:38
BLANK TEST	0.00	14:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:39
BLANK TEST	0.00	14:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:40
BLANK TEST	0.00	14:41

*Handwritten signature: M. Staw. C. H. 787*

