

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 11/14/2021

OPERATOR NAME:  
RYAN C SCHROECK  
Permit #: 41316  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:48
INTERNAL STANDARD	VERIFIED	20:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:49
BLANK TEST	0.00	20:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:50
BLANK TEST	0.00	20:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:51
BLANK TEST	0.00	20:52

