

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/24/2021

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:38
INTERNAL STANDARD	VERIFIED	12:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:39
BLANK TEST	0.00	12:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:40
BLANK TEST	0.00	12:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:41
BLANK TEST	0.00	12:42

Sgt. C. B. Geisler #24

