

SOLUTION CHANGE

WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 12/01/2021

OPERATOR NAME:  
 RYAN C SCHROECK  
 Permit #: 41316  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:40
INTERNAL STANDARD	VERIFIED	20:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:40
BLANK TEST	0.00	20:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:41
BLANK TEST	0.00	20:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:42
BLANK TEST	0.00	20:43

*RC 12/1/21*

