

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/10/2021

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:51
INTERNAL STANDARD	VERIFIED	21:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:51
BLANK TEST	0.00	21:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:53
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:54
BLANK TEST	0.00	21:55

Handwritten signature/initials

