

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/17/2021

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:50
INTERNAL STANDARD	VERIFIED	08:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:51
BLANK TEST	0.00	08:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:52
BLANK TEST	0.00	08:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:53
BLANK TEST	0.00	08:54



P. J. Bateholts 0714