

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/10/2022

OPERATOR NAME:
S T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:49
INTERNAL STANDARD	VERIFIED	11:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:49
BLANK TEST	0.00	11:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:52
BLANK TEST	0.00	11:52

Handwritten signature and initials

