

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 01/24/2022

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:09
INTERNAL STANDARD	VERIFIED	10:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12

*PO Steve Bunio #789*

