

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/01/2022

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:35
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37

Donnie J Long
944

