

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/08/2022

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:30
INTERNAL STANDARD	VERIFIED	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:33
BLANK TEST	0.00	09:33

PT Shw 44983

