

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/02/2022

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:08
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:12



PO Michael Campbell #157
PO M.A. CAMPBELL