

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/08/2022

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:44
INTERNAL STANDARD	VERIFIED	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:45
BLANK TEST	0.00	08:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48

PO. *[Signature]* #301

3/8/2022

