

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/15/2022



OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:41
INTERNAL STANDARD	VERIFIED	14:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:42
BLANK TEST	0.00	14:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	INTERFERENCE	14:43

Interference Detected

#1 *Don* *844*



SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/15/2022

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:46
INTERNAL STANDARD	VERIFIED	14:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:48
BLANK TEST	0.00	14:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:50

#2 *Don Long* 844

