

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/31/2022

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:53
INTERNAL STANDARD	VERIFIED	15:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:53
BLANK TEST	0.00	15:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:55
BLANK TEST	0.00	15:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:56
BLANK TEST	0.00	15:57



John D Bateholts 0714