

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 04/14/2022

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:41
INTERNAL STANDARD	VERIFIED	15:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:41
BLANK TEST	0.00	15:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44

*PO Steve Bunio #789*

